

Enforcement Notice summary

Name of provider:	Royal Cornwall Hospital NHS Trust
Location:	Royal Cornwall Hospital, 2 Penventinnie Lane, Treliske, Truro, Cornwall, TR1 3LQ
Notice Type:	Improvement Notice
Breach Date:	On and before 4 th November 2015
Breach:	Regulation 7 (1) and 7 (7)(b) Optimisation and 11(1) Training of the Ionising Radiation (Medical Exposure) Regulations 2000
Description:	An improvement notice was issued to the Trust under Section 21 of the Health and Safety at Work etc. Act 1974 following an inspection on 20 th April 2016 at Royal Cornwall Hospital. This inspection identified a radiographer who had not been adequately trained to carry out paediatric exposures in relation to a skeletal survey that they had been asked to carry out. In addition the trust had not optimised paediatric exposures since the installation of a new x-ray unit
Compliance deadline:	28 th June 2016
Result:	The Trust were inspected on the 11 th July 2016 and were deemed compliant with the improvement notice