

Enforcement Notice summary

Name of provider:	Birmingham Children's Hospital NHS Foundation Trust
Location:	Birmingham Children's Hospital, Steelhouse Lane, Birmingham, B4 6NH
Notice Type:	Improvement Notice
Breach Date:	29 th June 2016
Breach:	Regulation 11(4) of the Ionising Radiation (Medical Exposure) Regulations 2000
Description:	An improvement notice was issued to the Trust under Section 21 of the Health and Safety at Work etc. Act 1974 following an inspection on 29 th June 2016 at Birmingham Children's Hospital. This inspection identified a failure to record patient dose data for procedures performed using a mini C-arm and a failure to appropriately optimise exposures.
Compliance deadline:	12 th September 2016
Result:	The Trust were inspected on 5 th October 2016 and were deemed to be compliant with the improvement notice.