

Enforcement Notice summary

Name of provider:	Oxford University Hospitals NHS Foundation Trust
Location:	John Radcliffe Hospital Headley Way Headington Oxford OX3 9DU
Notice Type:	Improvement Notice
Breach Date:	On and before 18 th August 2016
Breach:	Regulation 4 (1)(a), 4(2), 4(3)(b), 4(4)(a)(b), 7(8), 8 and 11(1), 11(4) of the Ionising Radiation (Medical Exposure) Regulations 2000
Description:	An improvement notice was issued to the Trust under Section 21 of the Health and Safety at Work etc. Act 1974 following an inspection on 18th August 2016 at John Radcliffe Hospital. This inspection identified concerns around identification of non-medical referrers, quality assurance of procedures, the absence of a clinical audit programme and incomplete or absent training records for substantive and agency duty holders as required under IR(ME)R
Compliance deadline:	The Trust were inspected on 1 st February 2017 and were deemed to be compliant with the improvement notice.
Result:	Compliant