

## Enforcement Notice summary

<b>Name of provider:</b>	<b>Wye Valley NHS Trust</b>
<b>Location:</b>	Hereford County Hospital, Stonebow Road, Hereford, Herefordshire HR1 2ER
<b>Notice Type:</b>	Improvement Notice
<b>Breach Date:</b>	On and before 1st August 2018
<b>Breach:</b>	Regulation 8(3), 14(1), 15(6)(b)
<b>Description:</b>	An improvement notice was issued to the Trust under Section 21 of the Health and Safety at Work etc. Act 1974 following an inspection on 1 <sup>st</sup> August 2018 at the Hereford County Hospital. This inspection identified concerns in nuclear medicine on the system for recording analyses of accidental or unintended exposures, medical physics expert appointment and involvement, plus the specification of acceptable performance criteria for gamma camera equipment.
<b>Compliance deadline:</b>	<b>16<sup>th</sup> November 2018</b>
<b>Result:</b>	The trust was re-inspected on the 24 January 2019 and deemed to comply with this notice