

Ionising Radiation (Medical Exposure) Regulations 2017 Enforcement Notice summary – compliancy result

Name of employer:	Bradford Teaching Hospitals NHSFT
Location:	Bradford Royal Infirmary Bradford BD9 6RJ
Enforcement type:	Improvement Notice
Date notice served:	21 August 2019
Breach:	On and/or before 15 August 2019 The employer was found in breach of regulation/s 6(1)(a), 6(2) and 6(5)(b).
Description:	An improvement notice was issued to the employer under Section 21 of the Health and Safety at Work etc. Act 1974 following a planned IR(ME)R inspection of the radiology department. This inspection identified concerns surrounding the employer's procedures, which lacked sufficient detail to reflect clinical practice and were not subject to adequate quality assurance processes.
Result:	The employer was re-inspected on 16 December 2019 and deemed to comply with this notice.