



## **Ionising Radiation (Medical Exposure) Regulations 2017 Enforcement Notice summary**

**Name of Employer:** The Princess Alexandra Hospital NHS Trust

**Location:** The Princess Alexandra Hospital, Hamstel Rd, Harlow, CM20 1QZ

**Enforcement Type:** Improvement Notice

**Date Notice Served:** 10 May 2021

**Breach:** On and/or before 29 April 2021 the employer was found in breach of regulation 6(5)(c)

**Description:** An improvement notice was issued to the employer under section 21 of the Health and Safety at Work etc Act 1974 following an inspection of the radiology and cardiology departments. This inspection showed that the trust had not made available Diagnostic Reference Levels to operators working within the cardiology department.

**Compliance deadline:** 05 November 2021

**Result:** Ongoing